

BELGIUM	Radio Interface Specification	ULP-MMI	B14-04	V2.2 - 28/06/2011
----------------	--------------------------------------	----------------	---------------	--------------------------

	Nr	Parameter	description	Comments
Normative part	1	Radiocommunication Service	Mobile	
	2	Application	Ultra Low Power Medical Membrane Implants	The application is for Ultra Low Power medical membrane implants for blood pressure measurements.
	3	Frequency band	30-37.5 MHz	
	4	Channelling		
	5	Modulation/Occupied bandwidth		
	6	Direction/Separation		
	7	Transmit power/Power density	Max 1 mW ERP	
	8	Channel access and occupation rules	Duty Cycle < 10%	
	9	Authorisation regime	Licence exempt	Operation on NIB/NPB (Non interference basis/ Non protection basis)
	10	Additional essential requirements		
	11	Frequency planning assumption		
Informative part	12	Planned Change		
	13	Reference	EN 302 510-2 EC Decision 2006/771/EC ERC/REC 70-03	Only the version(s) mentioned in the most recent list of Harmonised Standards under the R&TTE directive as published by the European Commission in the OJEU can be used to enjoy presumption of conformity.
	14	Notification number		
	15	Remarks	Class 1 according Decision 2000/299/EC	